

Model E 28 | Drying and heating chambers Classic.Line with mechanical adjustment

BENEFITS

- Uniform drying conditions thanks to APT.line™ technology
- Identical test conditions throughout the chamber interior independent of sample size and quantity
- Outstanding thermal insulation saves operation costs



Model 28



Model 28

MAIN FEATURES

- Temperature range: 60 °C to 230 °C
- Adjustable exhaust air flap
- Hydro-mechanical thermostat
- Class 1 temperature limitation device (Modell Eo28-230V-T)
- Timer 0 – 120 min
- 2 chrome-plated racks

ORDERING INFORMATION

Interior volume [L]	Voltage	Option model	Version	Art.-No.
Model E 28				
28	230 V 1~ 50/60 Hz	Standard	Eo28-230V	9010-0001
		with TB class 1	Eo28-230V-T	9010-0003
	120 V 1~ 60 Hz	Standard	Eo28-120V	9010-0106

TECHNICAL DATA

Description	Eo28-230V	Eo28-230V-T	Eo28-120V
Article Number	9010-0001	9010-0003	9010-0106
Performance Data Temperature			
Temperature range [°C]	60...230	60...230	60...230
Heating-up time to 150 °C [min]	36	36	36
Recovery time after 30 seconds door open at 150 °C [min]	19	19	19
Electrical data			
Rated Voltage [V]	230	230	120
Power frequency [Hz]	50/60	50/60	60
Nominal power [kW]	0.8	0.8	0.8
Phase (Nominal voltage)	1~	1~	1~

Data Sheet Model E 28

Description	E028-230V	E028-230V-T	E028-120V
Article Number	9010-0001	9010-0003	9010-0106
Measures - Outer dimensions			
Width net [mm]	580	580	580
Height net [mm]	405	405	405
Depth net [mm]	425	425	425
Wall clearance back [mm]	160	160	160
Wall clearance sidewise [mm]	100	100	100
Measures - Doors			
Unit doors	1	1	1
Measures - Internal Dimensions			
Width [mm]	400	400	400
Height [mm]	280	280	280
Depth [mm]	250	250	250
Measures			
Interior volume [L]	28	28	28
Net weight of the unit (empty) [kg]	22	22	22
Interior volume [kg]	25	25	25
Load per rack [kg]	10	10	10

OPTIONS AND ACCESSORIES

Designation	Description	*	Art.-No.
Calibration certificate, expanded	each measurement at additional measuring point or testing temperature	-	8012-0022
Calibration certificate, temperature	for temperature, measurement in center of chamber at specified temperature	-	8012-0030
Magnetic Pouch (A4)	insert pouch with magnetic strip	-	1007-0065
pH-neutral detergent	concentrated, for gentle remove of residual contaminants; 1 kg	-	1002-0016
Rack	chrome plated	-	6004-0001
Rubber pads	set of 4 anti-slip feet, for safe stacking of unit	-	8012-0702
Shelf, perforated	stainless steel	-	6004-0028

* Notes › See last page

SERVICES

Designation	Description	*	Art.-No.
Installation services			
Installation	and set up of unit at operating location, connect to existing connections	13, 18	DL100100
Preventive maintenance			
Preventive maintenance	Executive of equipment inspection according to maintenance plan	14, 18	DL200100
Calibration services			
Calibration temperature	including certificate, one measuring point in center of chamber at specified temperature	14, 16, 17, 18	DL300101
Warranty service			
Extension of the warranty from 2 to 3 years	beginning with the date of delivery, wearing parts are not included	-	DL001041
Extension of the warranty from 2 to 5 years	beginning with the date of delivery, wearing parts are not included	-	DL001042

* Notes › See last page

NOTES

- 01 Condensation may occur in the area around the access port. Access ports may be placed in custom locations for an additional charge.
- 02 UL mark is not granted when this option is used.
- 03 Heat resistant only to max. 200 °C.
- 04 Only available on units rated for 230 V.
- 05 Attention: The pump is delivered separately. The electric connection for the pump (230 V/1~/50 Hz) must be carried out by an authorized electrician.
- 06 Heating-up time may increase as a result of the lower heat conductivity.
- 07 The additional heat input may affect the temperature behavior.
- 08 The pump (packaged separately) is ready for connection.
- 09 Not in conjunction with the optional access port, door with window and interior lighting.
- 10 Not available on 23-liter units.
- 11 Not available on 23- or 53-liter units.
- 12 Only available on units rated for 230 V or 400 V.
- 13 Installation and connections take place at unit location; transport within the company only upon consultation.
- 14 We recommend a BINDER service contract (see the chapter on BINDER Service) to cover unit inspections, calibrations and validations.
- 15 OQ according to Yellow Paper = completed factory validation documentation of all OQ checklists.
- 16 Sensor calibration is performed in an accredited calibration laboratory.
- 17 Calibration is performed according to the BINDER factory standard.
- 18 Quoted prices do not include travel costs. Please refer to the chapter on BINDER Service for travel costs for your region. Quoted prices for services performed in Switzerland do not include a country-specific added fee (available on request).
- 19 For additional accessories, refer to the Process documentation chapter.
- 20 When ordering IQ/OQ qualification folders and associated IQ/OQ execution on one order, we offer a 15 % discount for both items.
When ordering IQ/OQ/PQ qualification folders and associated IQ/OQ/PQ execution on one order, we offer a 15 % discount for the item of the IQ/OQ/PQ folder.
- 21 ATEX conformity:
Pumping chamber (pumped gases): II 2G IIC T3 X
Environment with inert purge gas: II 2G IIB T4 X
Environment without inert purge gas: II 3G IIB T4 X
Motor: II 2G Ex d IIB T4 Gb

BINDER GmbH
Tuttlingen, Germany
TEL +49 7462 2005 0
FAX +49 7462 2005 100
info@binder-world.com
www.binder-world.com

BINDER Asia Pacific (Hong Kong) Ltd.
Kowloon, Hong Kong, P.R. China
TEL +852 39070500
FAX +852 39070507
asia@binder-world.com
www.binder-world.com

Representative Office for CIS
Moscow, Russia
TEL +7 495 988 15 16
FAX +7 495 988 15 17
russia@binder-world.com
www.binder-world.ru

**BINDER Environmental Testing
Equipment (Shanghai) Co., Ltd.**
Shangha, P.R. China
TEL +86 21 685 808 25
FAX +86 21 685 808 29
china@binder-world.com
www.binder-world.com

BINDER Inc.
Bohemia, NY, USA
TEL +1 631 224 4340
FAX +1 631 224 4354
usa@binder-world.com
www.binder-world.us